

# BTL EMSELLA™

## CONTRAINDICATIONS – BTL EMSELLA

- Cardiac pacemakers
- implanted defibrillators, implanted neurostimulators
- electronic implants
- application in the area of growth plate
- pulmonary insufficiency
- metal implants
- drug pumps
- application in the head area
- hemorrhagic conditions
- anticoagulation therapy
- application in the heart area
- heart disorders
- malignant tumor
- fever
- pregnancy
- following recent surgical procedures when muscle contraction may disrupt the healing process
- application over menstruating uterus
- application over areas of the skin which lack normal sensation

The side effects may include, but are not limited to:

- muscular pain
- temporary muscle spasm
- temporary joint or tendon pain
- local erythema or skin redness